

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU052219\
 Data File : VU032254.D
 Acq On : 23 May 2019 05:08
 Operator : JC/SP
 Sample : K3001-17
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK02

Quant Time: May 23 07:51:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR052219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 23 07:22:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	364779	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	375259	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	151707	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	95346	4.65	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.00%
7) Chloroethane-d5	1.68	69	97649	5.11	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.20%
11) 1,1-Dichloroethene-d2	2.27	63	154303	4.09	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.80%
20) 2-Butanone-d5	4.20	46	292552	50.63	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.26%
24) Chloroform-d	4.64	84	260392	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.60%
26) 1,2-Dichloroethane-d4	5.31	65	143910	5.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.80%
32) Benzene-d6	5.33	84	502823	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.40%
36) 1,2-Dichloropropane-d6	6.33	67	167604	4.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.80%
41) Toluene-d8	7.56	98	455695	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.00%
43) trans-1,3-Dichloropropene-	7.85	79	60007	4.51	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.20%
46) 2-Hexanone-d5	8.31	63	246553	49.28	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.56%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	166196	4.84	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	212801	5.09	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU052219\
Data File : VU032254.D
Acq On : 23 May 2019 05:08
Operator : JC/SP
Sample : K3001-17
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK02

Quant Time: May 23 07:51:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR052219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 23 07:22:00 2019
Response via : Initial Calibration

