


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052220\
Data File : VU038277.D
Acq On : 22 May 2020 10:40
Operator : JC/MD
Sample : BFB29
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB29

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR050720WMA.M
Title : TRACE VOA SOM01.0
Last Update : Mon May 25 02:32:18 2020



AutoFind: Scans 2895, 2896, 2897; Background Corrected with Scan 2880

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	43594	PASS
75	95	30	80	51.8	118437	PASS
95	95	100	100	100.0	228693	PASS
96	95	5	9	6.6	15124	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	74.3	169962	PASS
175	174	5	9	7.4	12547	PASS
176	174	95	101	96.4	163904	PASS
177	176	5	9	6.4	10467	PASS

m/z	Abundance
39.80	319.0
41.20	100.0
44.00	1267.0

Instrument :
MSVOA_U
ClientSampleId :
BFB29

m/z	Abundance
36.00	1824.0
37.00	10054.0
38.00	9927.0
39.00	4058.0
39.90	605.0
42.90	129.0
44.00	1919.0
45.00	1932.0
46.00	355.0
47.10	2959.0
48.00	1432.0
49.00	9033.0
50.00	42568.0
51.00	14075.0
52.10	607.0
55.00	721.0
56.00	2995.0
57.00	6345.0
58.00	386.0
60.00	2128.0
61.00	10653.0
62.10	10666.0
63.10	7826.0
64.00	971.0
65.00	821.0
66.00	167.0
67.00	602.0
68.00	21720.0
69.10	24104.0
70.00	1567.0
72.10	1260.0
73.00	10412.0
74.00	39928.0
75.10	115040.0
76.10	10030.0
77.00	1072.0
77.90	963.0
79.00	6835.0
80.00	1647.0
80.90	5702.0
81.90	1539.0
82.80	107.0
85.60	164.0
87.00	9526.0
88.00	10018.0
90.90	862.0
92.00	6484.0
93.10	9344.0
94.10	26984.0
95.10	220800.0
96.10	15387.0
97.00	487.0
103.90	860.0
104.90	335.0
105.90	895.0
107.00	230.0
110.10	184.0
110.80	150.0
111.90	132.0
112.80	194.0
114.90	181.0
116.00	899.0
116.90	1323.0
118.00	704.0
119.00	1396.0
124.00	115.0
126.00	107.0
128.00	751.0
128.90	339.0
130.00	765.0
131.00	262.0
135.00	344.0
137.00	362.0
139.90	279.0
141.00	2154.0
141.90	366.0
143.00	2479.0
143.80	129.0
144.80	232.0
146.00	311.0
147.90	644.0
148.80	133.0
149.80	234.0
151.80	160.0
153.00	169.0
154.10	160.0
155.00	678.0
157.00	449.0
158.90	442.0
161.00	295.0
170.90	150.0

Instrument :
MSVOA_U
ClientSampleId :
BFB29

171.90	838.0
174.10	160128.0
175.10	12386.0
176.10	154176.0
177.00	10278.0
178.10	213.0

Instrument :
MSVOA_U
ClientSampleId :
BFB29

m/z	Abundance
36.10	2021.0
37.00	10799.0
38.00	10337.0
39.00	4039.0
39.90	570.0
43.00	136.0
44.00	2023.0
45.00	2169.0
47.10	3049.0
48.00	1554.0
49.00	9439.0
50.00	45040.0
51.00	14908.0
52.10	674.0
55.00	703.0
56.00	3335.0
57.00	6597.0
58.00	435.0
60.00	2364.0
61.00	10995.0
62.10	11060.0
63.10	8268.0
64.00	963.0
65.00	787.0
66.00	153.0
67.10	791.0
68.00	23264.0
69.00	25256.0
70.10	1703.0
72.10	1333.0
73.00	11061.0
74.00	42864.0
75.00	122328.0
76.10	10331.0
77.00	1112.0
77.90	933.0
79.00	7532.0
79.90	1645.0
80.90	6327.0
81.90	1542.0
85.70	210.0
87.00	10015.0
88.00	10246.0
90.90	993.0
92.00	6590.0
93.10	10149.0
94.10	28800.0
95.10	234752.0
96.10	15537.0
97.00	553.0
103.90	1010.0
104.90	378.0
106.00	1046.0
107.10	248.0
110.10	138.0
110.80	132.0
111.90	186.0
112.90	178.0
114.90	193.0
116.00	1043.0
116.90	1486.0
118.00	718.0
119.00	1553.0
128.00	891.0
128.90	392.0
130.00	690.0
131.00	295.0
135.00	411.0
136.90	393.0
139.30	110.0
139.90	274.0
141.00	2405.0
141.90	340.0
143.00	2658.0
143.80	152.0
144.90	295.0
146.00	320.0
148.00	607.0
148.80	154.0
149.80	271.0
151.90	130.0
153.00	213.0
154.10	205.0
155.00	692.0
156.90	497.0
158.90	453.0
161.00	271.0
170.80	152.0
172.00	926.0
174.10	174848.0
175.10	12861.0

Instrument :
MSVOA_U
ClientSampleId :
BFB29

176.10	168960.0
177.00	10851.0
178.00	268.0

Instrument :
MSVOA_U
ClientSampleId :
BFB29

m/z	Abundance
36.10	2015.0
37.00	10572.0
38.00	9892.0
39.00	3725.0
39.90	478.0
43.00	122.0
44.00	2102.0
45.00	2042.0
47.10	2843.0
48.00	1541.0
49.00	9000.0
50.00	43176.0
51.00	14126.0
52.10	585.0
53.00	139.0
55.00	601.0
56.00	3243.0
57.00	6129.0
58.00	417.0
60.00	2232.0
61.00	10947.0
62.10	10742.0
63.10	8143.0
64.00	938.0
64.90	619.0
67.10	810.0
68.00	23216.0
69.00	23928.0
70.10	1646.0
72.10	1263.0
73.00	10804.0
74.00	41088.0
75.00	117944.0
76.10	10222.0
77.00	1064.0
78.00	969.0
79.00	7244.0
79.90	1698.0
80.90	6301.0
81.90	1448.0
85.80	294.0
87.00	10039.0
88.00	9599.0
90.90	956.0
92.00	6161.0
93.10	9961.0
94.10	27504.0
95.10	230528.0
96.10	14448.0
96.90	529.0
103.90	1048.0
104.90	362.0
106.00	1001.0
107.10	254.0
110.90	157.0
112.00	217.0
112.90	134.0
114.90	178.0
116.00	1046.0
116.90	1628.0
118.00	761.0
119.00	1473.0
124.00	105.0
128.00	927.0
128.90	435.0
129.90	587.0
131.00	277.0
135.00	435.0
136.90	378.0
140.00	254.0
141.00	2347.0
141.90	253.0
143.00	2519.0
143.90	119.0
144.90	318.0
146.00	309.0
147.00	132.0
148.00	563.0
148.80	149.0
149.90	268.0
153.00	175.0
154.10	193.0
155.00	598.0
156.30	126.0
156.90	456.0
157.80	101.0
158.90	391.0
161.00	240.0
170.70	124.0
172.00	921.0
174.00	174912.0

Instrument :
MSVOA_U
ClientSampleId :
BFB29

175.10	12394.0
176.10	168576.0
177.00	10272.0
178.00	340.0

Instrument :
MSVOA_U
ClientSampleId :
BFB29