

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052223\
 Data File : VU054248.D
 Acq On : 22 May 2023 16:44
 Operator : JC/MD
 Sample : 02874-05DL 50X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0NY7DL

Quant Time: May 23 04:01:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue May 23 02:58:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	145926	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	139632	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	66380	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	39934	5.076	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	101.600%	
7) Chloroethane-d5	1.912	69	33266	5.652	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	113.000%	
11) 1,1-Dichloroethene-d2	2.562	65	18869	5.310	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	106.200%	
20) 2-Butanone-d5	4.632	46	91729	65.380	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	130.760%#	
24) Chloroform-d	5.057	84	79586	5.381	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	107.600%	
26) 1,2-Dichloroethane-d4	5.697	65	38985	5.729	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	114.600%	
32) Benzene-d6	5.722	84	157918	5.554	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	111.000%	
36) 1,2-Dichloropropane-d6	6.684	67	48725	5.928	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	118.600%	
41) Toluene-d8	7.893	98	132798	4.946	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	99.000%	
43) trans-1,3-Dichloroprop...	8.176	79	17603	5.303	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	106.000%	
46) 2-Hexanone-d5	8.629	63	63831	64.580	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	129.160%	
56) 1,1,2,2-Tetrachloroeth...	10.751	84	35904	6.095	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	121.800%#	
66) 1,2-Dichlorobenzene-d4	12.188	152	48966	6.045	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	121.000%#	
Target Compounds						Qvalue
16) Methylene chloride	3.041	84	26192	2.240	ug/L	96
22) cis-1,2-Dichloroethene	4.661	96	12413	1.169	ug/L	95
34) Trichloroethene	6.536	95	106644	10.033	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052223\
Data File : VU054248.D
Acq On : 22 May 2023 16:44
Operator : JC/MD
Sample : 02874-05DL 50X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0NY7DL

Quant Time: May 23 04:01:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue May 23 02:58:02 2023
Response via : Initial Calibration

