

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052223\
 Data File : VU054248.D
 Acq On : 22 May 2023 16:44
 Operator : JC/MD
 Sample : 02874-05DL 50X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0NY7DL

Quant Time: May 23 04:01:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue May 23 02:58:02 2023
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards							
1) 1,4-Difluorobenzene		6.243	114	145926	5.000	ug/L	0.00
28) Chlorobenzene-d5		9.414	117	139632	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.806	152	66380	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.597	65	39934	5.076	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.600%	
7) Chloroethane-d5		1.912	69	33266	5.652	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.000%	
11) 1,1-Dichloroethene-d2		2.562	65	18869	5.310	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	106.200%	
20) 2-Butanone-d5		4.632	46	91729	65.380	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	130.760%#	
24) Chloroform-d		5.057	84	79586	5.381	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.600%	
26) 1,2-Dichloroethane-d4		5.697	65	38985	5.729	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.600%	
32) Benzene-d6		5.722	84	157918	5.554	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.000%	
36) 1,2-Dichloropropane-d6		6.684	67	48725	5.928	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	118.600%	
41) Toluene-d8		7.893	98	132798	4.946	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.000%	
43) trans-1,3-Dichloroprop...		8.176	79	17603	5.303	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	106.000%	
46) 2-Hexanone-d5		8.629	63	63831	64.580	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	129.160%	
56) 1,1,2,2-Tetrachloroeth...		10.751	84	35904	6.095	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	121.800%#	
66) 1,2-Dichlorobenzene-d4		12.188	152	48966	6.045	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	121.000%#	
Target Compounds							
16) Methylene chloride		3.041	84	26192	2.240	ug/L	96
22) cis-1,2-Dichloroethene		4.661	96	12413	1.169	ug/L	95
34) Trichloroethene		6.536	95	106644	10.033	ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052223\
Data File : VU054248.D
Acq On : 22 May 2023 16:44
Operator : JC/MD
Sample : 02874-05DL 50X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0NY7DL

Quant Time: May 23 04:01:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue May 23 02:58:02 2023
Response via : Initial Calibration









#16
Methylene chloride
Concen: 2.240 ug/L
RT: 3.041 min Scan# 51
Delta R.T. -0.000 min
Lab File: VU054248.D
Acq: 22 May 2023 16:44




Instrument :
MSVOA_U
ClientSampleId :
CONY7DL


Tgt Ion: 84 Resp: 26192
Ion Ratio Lower Upper
84 100
86 60.2 45.5 84.5
49 113.9 81.3 150.9



#22
cis-1,2-Dichloroethene
Concen: 1.169 ug/L
RT: 4.661 min Scan# 1033
Delta R.T. -0.000 min
Lab File: VU054248.D
Acq: 22 May 2023 16:44


Tgt Ion: 96 Resp: 12413
Ion Ratio Lower Upper
96 100
61 118.4 85.0 157.8
98 58.7 45.8 85.0





#34
 Trichloroethene
 Concen: 10.033 ug/L
 RT: 6.536 min Scan# 1
 Delta R.T. -0.003 min
 Lab File: VU054248.D
 Acq: 22 May 2023 16:44

Instrument :
 MSVOA_U
 ClientSampleId :
 C0NY7DL



Tgt Ion: 95 Resp: 106644

Ion	Ratio	Lower	Upper
95	100		
97	64.7	45.8	85.0
132	98.7	70.1	130.3
130	105.6	72.4	134.5

