

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052421\
 Data File : VU043921.D
 Acq On : 24 May 2021 20:25
 Operator : SY/MD
 Sample : M2494-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GB420

Quant Time: May 25 04:13:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR052421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue May 25 04:12:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	169695	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	173132	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	75457	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	60065	6.36	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	127.20%
7) Chloroethane-d5	1.922	69	43014	6.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	126.20%
11) 1,1-Dichloroethene-d2	2.578	65	26980	5.87	ug/L	0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	117.40%
20) 2-Butanone-d5	4.649	46	124532	41.58	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	83.16%
24) Chloroform-d	5.076	84	124691	5.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	114.40%
26) 1,2-Dichloroethane-d4	5.713	65	64496	5.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	115.60%
32) Benzene-d6	5.739	84	252252	5.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	114.80%
36) 1,2-Dichloropropane-d6	6.700	67	81453	5.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	116.40%
41) Toluene-d8	7.906	98	210646	5.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.20%
43) trans-1,3-Dichloroprop...	8.185	79	28616	5.36	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	107.20%
46) 2-Hexanone-d5	8.648	63	129615	55.38	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.76%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	72806	5.66	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	113.20%
66) 1,2-Dichlorobenzene-d4	12.201	152	85323	6.22	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	124.40%#
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.591	96	40832	3.72	ug/L	# 85
13) Acetone	2.652	43	3598	2.28	ug/L	100
19) 1,1-Dichloroethane	3.883	63	19139	0.98	ug/L	97
29) 1,1,1-Trichloroethane	5.327	97	459315	25.61	ug/L	99
34) Trichloroethene	6.552	95	1202	0.10	ug/L	95
47) Tetrachloroethene	8.558	164	89745	8.55	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052421\
Data File : VU043921.D
Acq On : 24 May 2021 20:25
Operator : SY/MD
Sample : M2494-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GB420

Quant Time: May 25 04:13:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR052421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue May 25 04:12:21 2021
Response via : Initial Calibration

