

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052421\
 Data File : VU043949.D
 Acq On : 25 May 2021 09:11
 Operator : SY/MD
 Sample : VU0524WBL03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK155

Quant Time: May 25 14:28:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR052421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue May 25 07:46:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	207599	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	216103	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	92417	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	54484	4.72	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.40%
7) Chloroethane-d5	1.912	69	41520	4.98	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.60%
11) 1,1-Dichloroethene-d2	2.565	65	22389	3.98	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.60%
20) 2-Butanone-d5	4.703	46	211418	57.70	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	115.40%
24) Chloroform-d	5.073	84	137232	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.80%
26) 1,2-Dichloroethane-d4	5.716	65	73325	5.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.40%
32) Benzene-d6	5.732	84	272602	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.40%
36) 1,2-Dichloropropane-d6	6.700	67	88195	5.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.00%
41) Toluene-d8	7.906	98	223703	4.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.60%
43) trans-1,3-Dichloroprop...	8.189	79	32222	4.83	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.60%
46) 2-Hexanone-d5	8.652	63	152580	52.23	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.46%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	82796	5.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.20%
66) 1,2-Dichlorobenzene-d4	12.201	152	91428	5.44	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.80%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052421\
Data File : VU043949.D
Acq On : 25 May 2021 09:11
Operator : SY/MD
Sample : VU0524WBL03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK155

Quant Time: May 25 14:28:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR052421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue May 25 07:46:23 2021
Response via : Initial Calibration

