

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052424\
 Data File : VU058978.D
 Acq On : 24 May 2024 18:10
 Operator : MD/SY
 Sample : P2568-06ME
 Misc : 6.03g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH5Q7ME

Quant Time: May 24 22:30:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052324WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 24 22:25:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	195049	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	198834	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	83082	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	78494	38.050	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	76.100%		
7) Chloroethane-d5	1.872	69	51683	30.146	ug/L	-0.04
Spiked Amount 50.000	Range 70 - 130		Recovery =	60.300%#		
11) 1,1-Dichloroethene-d2	2.531	63	105809	33.572	ug/L	-0.03
Spiked Amount 50.000	Range 60 - 125		Recovery =	67.140%		
21) 2-Butanone-d5	4.624	46	137951	97.353	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	97.350%		
24) Chloroform-d	5.049	84	141059	39.865	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	79.720%		
26) 1,2-Dichloroethane-d4	5.689	65	95754	44.248	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	88.500%		
32) Benzene-d6	5.714	84	295551	43.600	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	87.200%		
36) 1,2-Dichloropropane-d6	6.682	67	99854	44.598	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	89.200%		
41) Toluene-d8	7.891	98	248423	40.886	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	81.780%		
43) trans-1,3-Dichloroprop...	8.174	79	38860	38.653	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	77.300%		
47) 2-Hexanone-d5	8.640	63	85723	92.573	ug/L	0.01
Spiked Amount 100.000	Range 45 - 130		Recovery =	92.570%		
56) 1,1,2,2-Tetrachloroeth...	10.756	84	145014	43.052	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	86.100%		
66) 1,2-Dichlorobenzene-d4	12.193	152	83845	48.909	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	97.820%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052424\
Data File : VU058978.D
Acq On : 24 May 2024 18:10
Operator : MD/SY
Sample : P2568-06ME
Misc : 6.03g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH5Q7ME

Quant Time: May 24 22:30:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052324WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 24 22:25:41 2024
Response via : Initial Calibration

