

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052424\
 Data File : VU058983.D
 Acq On : 24 May 2024 20:13
 Operator : MD/SY
 Sample : P2566-13 1000X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHHS0

Quant Time: May 24 22:31:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052324WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 24 22:25:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	168967	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	164697	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	67071	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	78147	43.729	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.460%
7) Chloroethane-d5	1.885	69	67290	45.307	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.620%
11) 1,1-Dichloroethene-d2	2.554	63	97119	35.571	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.140%
21) 2-Butanone-d5	4.608	46	146580	119.410	ug/L	-0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	119.410%
24) Chloroform-d	5.052	84	145883	47.592	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.180%
26) 1,2-Dichloroethane-d4	5.692	65	95779	51.092	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.180%
32) Benzene-d6	5.718	84	289154	51.498	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.000%
36) 1,2-Dichloropropane-d6	6.682	67	100387	54.130	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.260%
41) Toluene-d8	7.891	98	229779	45.656	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.320%
43) trans-1,3-Dichloroprop...	8.174	79	37680	45.248	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.500%
47) 2-Hexanone-d5	8.624	63	89262	116.375	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	116.370%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	149003	53.406	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	106.820%
66) 1,2-Dichlorobenzene-d4	12.187	152	75628	54.647	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052424\
Data File : VU058983.D
Acq On : 24 May 2024 20:13
Operator : MD/SY
Sample : P2566-13 1000X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHHS0

Quant Time: May 24 22:31:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052324WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 24 22:25:41 2024
Response via : Initial Calibration

