

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052521\
 Data File : VU043967.D
 Acq On : 25 May 2021 17:32
 Operator : SY/MD
 Sample : VSTD0.136
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.1136

Quant Time: May 26 04:08:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052521.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 26 04:08:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.207	114	16463	0.50	ug/L	-0.02
5) Chlorobenzene-d5	9.361	117	14635	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.742	152	6652	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.592	65	702	0.06	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.680	65	1798	0.11	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.639	67	1395	0.10	ug/L	-0.02
8) Toluene-d8	7.845	98	3226	0.09	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.679	84	1309	0.11	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.599	62	1376	0.10	ug/L #	43
6) Trichloroethene	6.508	95	1378	0.10	ug/L	98
9) 1,2-Dibromoethane	8.866	107	990	0.11	ug/L #	98
12) 1,2,3-Trichloropropane	10.748	75	913	0.12	ug/L	97
13) 1,2-Dibromo-3-chloropr...	12.924	75	186	0.13	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052521\
Data File : VU043967.D
Acq On : 25 May 2021 17:32
Operator : SY/MD
Sample : VSTD0.136
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.1136

Quant Time: May 26 04:08:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052521.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 26 04:08:16 2021
Response via : Initial Calibration

