

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052521\
 Data File : VU043971.D
 Acq On : 25 May 2021 19:08
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VICV134

Quant Time: May 26 04:30:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052521.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 26 04:28:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.225	114	16650	0.50	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	15270	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.759	152	6988	0.50	ug/L	0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.606	65	4278	0.54	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 108.00%			
4) 1,2-Dichloroethane-d4	5.680	65	6752	0.43	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 86.00%			
7) 1,2-Dichloropropane-d6	6.658	67	7937	0.53	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 106.00%			
8) Toluene-d8	7.845	98	15583	0.47	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 94.00%			
10) 1,1,2,2-Tetrachloroeth...	10.696	84	7452	0.52	ug/L	0.02
Spiked Amount 0.500	Range 65 - 120		Recovery = 104.00%			
Target Compounds						
						Qvalue
3) Vinyl chloride	1.606	62	7151	0.48	ug/L	99
6) Trichloroethene	6.508	95	7169	0.48	ug/L	100
9) 1,2-Dibromoethane	8.866	107	5264	0.49	ug/L	99
12) 1,2,3-Trichloropropane	10.748	75	4967	0.50	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.924	75	961	0.48	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052521\
Data File : VU043971.D
Acq On : 25 May 2021 19:08
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VICV134

Quant Time: May 26 04:30:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052521.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 26 04:28:20 2021
Response via : Initial Calibration

