

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052521\
 Data File : VU043966.D
 Acq On : 25 May 2021 16:44
 Operator : SY/MD
 Sample : VSTD0.0535
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.05135

Quant Time: May 26 04:08:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052521.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 26 04:08:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.225	114	17415	0.50	ug/L	0.00
5) Chlorobenzene-d5	9.365	117	15490	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.745	152	7451	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.592	65	328	0.03	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.680	65	1154	0.07	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.658	67	665	0.05	ug/L	0.00
8) Toluene-d8	7.849	98	1983	0.05	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.683	84	712	0.06	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.599	62	795	0.06	ug/L #	1
6) Trichloroethene	6.508	95	811	0.06	ug/L	97
9) 1,2-Dibromoethane	8.870	107	559	0.06	ug/L #	95
12) 1,2,3-Trichloropropane	10.751	75	544	0.06	ug/L	97
13) 1,2-Dibromo-3-chloropr...	12.928	75	128	0.08	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052521\
Data File : VU043966.D
Acq On : 25 May 2021 16:44
Operator : SY/MD
Sample : VSTD0.0535
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.05135

Quant Time: May 26 04:08:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052521.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 26 04:08:16 2021
Response via : Initial Calibration

