

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052521\
 Data File : VU043968.D
 Acq On : 25 May 2021 17:56
 Operator : SY/MD
 Sample : VSTD0.531
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5131

Quant Time: May 26 04:08:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052521.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 26 04:08:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.225	114	16001	0.50	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	14728	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.742	152	6759	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.599	65	4029	0.35	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.680	65	6706	0.42	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.658	67	7859	0.59	ug/L	0.00
8) Toluene-d8	7.845	98	15232	0.43	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.679	84	7369	0.64	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.599	62	6955	0.53	ug/L	100
6) Trichloroethene	6.508	95	7266	0.54	ug/L	100
9) 1,2-Dibromoethane	8.866	107	5358	0.58	ug/L	100
12) 1,2,3-Trichloropropane	10.748	75	4941	0.64	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.924	75	941	0.63	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052521\
Data File : VU043968.D
Acq On : 25 May 2021 17:56
Operator : SY/MD
Sample : VSTD0.531
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5131

Quant Time: May 26 04:08:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052521.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 26 04:08:16 2021
Response via : Initial Calibration

