

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052521\
 Data File : VU043969.D
 Acq On : 25 May 2021 18:20
 Operator : SY/MD
 Sample : VSTD00132
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD001132

Quant Time: May 26 04:09:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052521.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 26 04:08:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.225	114	16732	0.50	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	15525	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.758	152	7082	0.50	ug/L	0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.605	65	9371	0.78	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.679	65	13043	0.78	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.658	67	16113	1.14	ug/L	0.00
8) Toluene-d8	7.845	98	31568	0.84	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.696	84	14823	1.21	ug/L	0.02
Target Compounds						
						Qvalue
3) Vinyl chloride	1.605	62	16040	1.17	ug/L	98
6) Trichloroethene	6.507	95	14825	1.05	ug/L	100
9) 1,2-Dibromoethane	8.866	107	10863	1.12	ug/L #	100
12) 1,2,3-Trichloropropane	10.747	75	9997	1.23	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.924	75	1884	1.21	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052521\
Data File : VU043969.D
Acq On : 25 May 2021 18:20
Operator : SY/MD
Sample : VSTD00132
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD001132

Quant Time: May 26 04:09:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052521.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 26 04:08:16 2021
Response via : Initial Calibration

