

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052522\
 Data File : VU048662.D
 Acq On : 25 May 2022 11:36
 Operator : SY/MD
 Sample : VU0525MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK137

Quant Time: May 26 06:51:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu May 26 06:51:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	202169	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	205758	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	112793	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	40242	32.499	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	65.000%		
7) Chloroethane-d5	1.912	69	36648	39.504	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	79.000%		
11) 1,1-Dichloroethene-d2	2.568	63	81068	28.948	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	57.900%#		
21) 2-Butanone-d5	4.617	46	115024	97.613	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	97.610%		
24) Chloroform-d	5.063	84	145835	46.606	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	93.220%		
26) 1,2-Dichloroethane-d4	5.703	65	107992	48.920	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	97.840%		
32) Benzene-d6	5.729	84	235775	41.173	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	82.340%		
36) 1,2-Dichloropropane-d6	6.690	67	71431	42.491	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	84.980%		
41) Toluene-d8	7.899	98	214667	39.854	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	79.700%#		
43) trans-1,3-Dichloroprop...	8.179	79	42567	39.598	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	79.200%		
47) 2-Hexanone-d5	8.632	63	79658	86.403	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	86.400%		
56) 1,1,2,2-Tetrachloroeth...	10.755	84	127312	45.767	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	91.540%		
66) 1,2-Dichlorobenzene-d4	12.195	152	107920	46.803	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	93.600%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052522\
Data File : VU048662.D
Acq On : 25 May 2022 11:36
Operator : SY/MD
Sample : VU0525MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK137

Quant Time: May 26 06:51:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
Quant Title : VOC Analysis
QLast Update : Thu May 26 06:51:07 2022
Response via : Initial Calibration

