

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052621\
 Data File : VU043976.D
 Acq On : 26 May 2021 11:20
 Operator : SY/MD
 Sample : VSTD0.0536
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.05136

Quant Time: May 27 01:28:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052621.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 27 01:26:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.207	114	16165	0.50	ug/L	-0.02
5) Chlorobenzene-d5	9.365	117	14329	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.745	152	6501	0.50	ug/L	-0.01
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.592	65	390	0.04	ug/L	-0.01
4) 1,2-Dichloroethane-d4	5.680	65	1189	0.08	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.639	67	679	0.05	ug/L	-0.02
8) Toluene-d8	7.849	98	1774	0.06	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.682	84	647	0.05	ug/L	-0.01
Target Compounds						Qvalue
3) Vinyl chloride	1.599	62	675	0.04	ug/L #	1
6) Trichloroethene	6.508	95	591	0.05	ug/L	97
9) 1,2-Dibromoethane	8.870	107	401	0.04	ug/L #	93
12) 1,2,3-Trichloropropane	10.751	75	379	0.04	ug/L	96
13) 1,2-Dibromo-3-chloropr...	12.928	75	98	0.05	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052621\
Data File : VU043976.D
Acq On : 26 May 2021 11:20
Operator : SY/MD
Sample : VSTD0.0536
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.05136

Quant Time: May 27 01:28:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052621.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 27 01:26:40 2021
Response via : Initial Calibration

