

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052621\
 Data File : VU043978.D
 Acq On : 26 May 2021 12:12
 Operator : SY/MD
 Sample : VSTD0.532
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5132

Quant Time: May 27 01:28:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052621.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 27 01:26:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.225	114	15708	0.50	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	14365	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.759	152	6384	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.605	65	4973	0.52	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.680	65	6987	0.49	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.658	67	8012	0.58	ug/L	0.00
8) Toluene-d8	7.845	98	15318	0.54	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.696	84	7284	0.55	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.605	62	8773	0.52	ug/L	100
6) Trichloroethene	6.508	95	7312	0.56	ug/L	100
9) 1,2-Dibromoethane	8.866	107	5311	0.57	ug/L	100
12) 1,2,3-Trichloropropane	10.748	75	4994	0.52	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.924	75	960	0.54	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052621\
Data File : VU043978.D
Acq On : 26 May 2021 12:12
Operator : SY/MD
Sample : VSTD0.532
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5132

Quant Time: May 27 01:28:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052621.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 27 01:26:40 2021
Response via : Initial Calibration

