

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052621\
 Data File : VU043979.D
 Acq On : 26 May 2021 12:36
 Operator : SY/MD
 Sample : VSTD00133
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD001133

Quant Time: May 27 01:28:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052621.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 27 01:26:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.225	114	16699	0.50	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	15429	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.759	152	6907	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.605	65	9490	0.93	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.679	65	13393	0.88	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.658	67	16279	1.10	ug/L	0.00
8) Toluene-d8	7.845	98	31417	1.03	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.696	84	14972	1.06	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.605	62	15924	0.88	ug/L	98
6) Trichloroethene	6.507	95	14555	1.03	ug/L	99
9) 1,2-Dibromoethane	8.866	107	10548	1.05	ug/L #	100
12) 1,2,3-Trichloropropane	10.748	75	10084	0.97	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.924	75	1890	0.98	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052621\
Data File : VU043979.D
Acq On : 26 May 2021 12:36
Operator : SY/MD
Sample : VSTD00133
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD001133

Quant Time: May 27 01:28:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052621.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 27 01:26:40 2021
Response via : Initial Calibration

