

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052621\
 Data File : VU043981.D
 Acq On : 26 May 2021 13:24
 Operator : SY/MD
 Sample : VIBLK311
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK311

Quant Time: May 27 01:37:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052621.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 27 01:35:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.225	114	15366	0.50	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	13806	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.759	152	6230	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.598	65	3821	0.41	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	82.00%
4) 1,2-Dichloroethane-d4	5.679	65	6544	0.47	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	94.00%
7) 1,2-Dichloropropane-d6	6.658	67	7610	0.57	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	114.00%
8) Toluene-d8	7.845	98	14802	0.54	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	108.00%
10) 1,1,2,2-Tetrachloroeth...	10.696	84	7076	0.56	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	112.00%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052621\
Data File : VU043981.D
Acq On : 26 May 2021 13:24
Operator : SY/MD
Sample : VIBLK311
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK311

Quant Time: May 27 01:37:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052621.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 27 01:35:45 2021
Response via : Initial Calibration

