

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052621\
 Data File : VU043983.D
 Acq On : 26 May 2021 16:53
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VICV135

Quant Time: May 27 01:37:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052621.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 27 01:35:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.225	114	15205	0.50	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	14145	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.759	152	6494	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.606	65	4943	0.54	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 108.00%			
4) 1,2-Dichloroethane-d4	5.680	65	6766	0.49	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 98.00%			
7) 1,2-Dichloropropane-d6	6.658	67	7906	0.58	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 116.00%			
8) Toluene-d8	7.845	98	15219	0.54	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 108.00%			
10) 1,1,2,2-Tetrachloroeth...	10.696	84	7806	0.60	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 120.00%			
Target Compounds						
						Qvalue
3) Vinyl chloride	1.606	62	8238	0.50	ug/L	98
6) Trichloroethene	6.508	95	7048	0.55	ug/L	99
9) 1,2-Dibromoethane	8.866	107	5394	0.58	ug/L	99
12) 1,2,3-Trichloropropane	10.748	75	5224	0.54	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.924	75	924	0.51	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052621\
Data File : VU043983.D
Acq On : 26 May 2021 16:53
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VICV135

Quant Time: May 27 01:37:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052621.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 27 01:35:45 2021
Response via : Initial Calibration

