

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052621\
 Data File : VU043980.D
 Acq On : 26 May 2021 13:00
 Operator : SY/MD
 Sample : VSTD00234
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD002134

Quant Time: May 27 01:28:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052621.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 27 01:26:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Difluorobenzene	6.225	114	17456	0.50 ug/L	0.00
5) Chlorobenzene-d5	9.364	117	16181	0.50 ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.745	152	7132	0.50 ug/L	-0.01
System Monitoring Compounds					
2) Vinyl Chloride-d3	1.605	65	19234	1.81 ug/L	0.00
4) 1,2-Dichloroethane-d4	5.679	65	25747	1.61 ug/L	0.00
7) 1,2-Dichloropropane-d6	6.658	67	32306	2.08 ug/L	0.00
8) Toluene-d8	7.849	98	64786	2.02 ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.682	84	29588	2.00 ug/L	-0.01
Target Compounds					
3) Vinyl chloride	1.605	62	37803	2.00 ug/L	Qvalue 96
6) Trichloroethene	6.508	95	33905	2.30 ug/L	99
9) 1,2-Dibromoethane	8.869	107	24440	2.31 ug/L #	98
12) 1,2,3-Trichloropropane	10.751	75	23153	2.16 ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.927	75	4379	2.20 ug/L	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052621\
Data File : VU043980.D
Acq On : 26 May 2021 13:00
Operator : SY/MD
Sample : VSTD00234
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD002134

Quant Time: May 27 01:28:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052621.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 27 01:26:40 2021
Response via : Initial Calibration

