


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052622\
 Data File : VU048668.D
 Acq On : 26 May 2022 09:20
 Operator : SY/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Title : SW846 8260
 Last Update : Fri May 27 06:42:31 2022



AutoFind: Scans 2890, 2891, 2892; Background Corrected with Scan 2877

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.6	6522	PASS
75	95	30	60	55.3	20485	PASS
95	95	100	100	100.0	37040	PASS
96	95	5	9	8.1	2992	PASS
173	174	0.00	2	0.4	148	PASS
174	95	50	100	96.2	35632	PASS
175	174	5	9	7.2	2559	PASS
176	174	95	101	97.6	34779	PASS
177	176	5	9	6.4	2215	PASS