

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052623\
 Data File : VU054275.D
 Acq On : 26 May 2023 11:01
 Operator : JC/MD
 Sample : VU0526WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0526WBL01

Quant Time: May 29 23:58:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.366	168	253950	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	404831	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.414	117	390813	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.809	152	198494	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	168776	50.698	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.400%
35) Dibromofluoromethane	5.282	113	136062	52.191	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.380%
50) Toluene-d8	7.893	98	507559	51.184	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.360%
62) 4-Bromofluorobenzene	10.626	95	182411	47.441	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052623\
Data File : VU054275.D
Acq On : 26 May 2023 11:01
Operator : JC/MD
Sample : VU0526WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0526WBL01

Quant Time: May 29 23:58:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

