

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052623\
 Data File : VU054281.D
 Acq On : 26 May 2023 13:28
 Operator : JC/MD
 Sample : 02936-04 50X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BRP2-20230523

Quant Time: May 30 00:00:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.366	168	256781	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	408465	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.414	117	388257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	193941	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	164257	48.797	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.600%		
35) Dibromofluoromethane	5.282	113	133872	50.894	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	101.780%		
50) Toluene-d8	7.893	98	497551	49.729	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.460%		
62) 4-Bromofluorobenzene	10.626	95	181017	46.660	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	93.320%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.382	85	1874	0.654	ug/l	96
4) Vinyl Chloride	1.604	62	72481	20.239	ug/l	99
12) 1,1-Dichloroethene	2.578	96	726	0.263	ug/l	96
21) trans-1,2-Dichloroethene	3.350	96	6663	2.253	ug/l	93
27) cis-1,2-Dichloroethene	4.658	96	277174	84.259	ug/l	93
44) Trichloroethene	6.539	130	37492	11.381	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052623\
Data File : VU054281.D
Acq On : 26 May 2023 13:28
Operator : JC/MD
Sample : 02936-04 50X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BRP2-20230523

Quant Time: May 30 00:00:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

