

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052623\
 Data File : VU054286.D
 Acq On : 26 May 2023 15:26
 Operator : JC/MD
 Sample : 02936-20
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FB-01-20230524

Quant Time: May 30 00:02:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.366	168	253206	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.244	114	401820	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.411	117	390281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	196467	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	166540	50.174	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.340%
35) Dibromofluoromethane	5.282	113	136726	52.838	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.680%
50) Toluene-d8	7.893	98	496766	50.471	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.940%
62) 4-Bromofluorobenzene	10.626	95	180773	47.368	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.740%
Target Compounds						
					Qvalue	
16) Acetone	2.636	43	7979	4.194	ug/l #	83
20) Methylene Chloride	3.038	84	1401	0.408	ug/l #	82
30) Chloroform	5.083	83	59244	10.874	ug/l	96
44) Trichloroethene	6.539	130	3599	1.111	ug/l	90
47) Bromodichloromethane	7.099	83	6867	1.649	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052623\
Data File : VU054286.D
Acq On : 26 May 2023 15:26
Operator : JC/MD
Sample : 02936-20
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FB-01-20230524

Quant Time: May 30 00:02:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

