

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052623\
 Data File : VU054289.D
 Acq On : 26 May 2023 16:37
 Operator : JC/MD
 Sample : 02936-11DL 5X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 8BR-20230524DL

Quant Time: May 30 00:03:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	259231	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	418868	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.410	117	381942	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	187647	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	171299	50.408	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.820%	
35) Dibromofluoromethane	5.282	113	138593	51.380	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.760%	
50) Toluene-d8	7.893	98	513139	50.013	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.020%	
62) 4-Bromofluorobenzene	10.629	95	174402	43.838	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 87.680%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.571	101	2486	0.923	ug/l	96
21) trans-1,2-Dichloroethene	3.353	96	1258	0.421	ug/l	83
27) cis-1,2-Dichloroethene	4.661	96	8931	2.689	ug/l	94
44) Trichloroethene	6.536	130	143921	42.603	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052623\
Data File : VU054289.D
Acq On : 26 May 2023 16:37
Operator : JC/MD
Sample : 02936-11DL 5X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
8BR-20230524DL

Quant Time: May 30 00:03:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

