

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052623\
 Data File : VU054291.D
 Acq On : 26 May 2023 17:24
 Operator : JC/MD
 Sample : 02936-13 100X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 7BR-20230524

Quant Time: May 30 00:04:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.366	168	252773	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.244	114	402059	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.411	117	375015	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	187348	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	163782	49.427	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.860%		
35) Dibromofluoromethane	5.282	113	133486	51.556	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	103.120%		
50) Toluene-d8	7.893	98	489210	49.674	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.340%		
62) 4-Bromofluorobenzene	10.626	95	170840	44.738	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	89.480%		
Target Compounds						Qvalue
4) Vinyl Chloride	1.604	62	13728	3.894	ug/l	96
12) 1,1-Dichloroethene	2.578	96	705	0.259	ug/l	90
20) Methylene Chloride	3.044	84	2310	0.674	ug/l	96
21) trans-1,2-Dichloroethene	3.350	96	2293	0.788	ug/l	95
27) cis-1,2-Dichloroethene	4.658	96	223743	69.094	ug/l	95
44) Trichloroethene	6.536	130	23581	7.272	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052623\
Data File : VU054291.D
Acq On : 26 May 2023 17:24
Operator : JC/MD
Sample : 02936-13 100X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
7BR-20230524

Quant Time: May 30 00:04:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

