

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052623\
 Data File : VU054293.D
 Acq On : 26 May 2023 18:12
 Operator : JC/MD
 Sample : 02936-15
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 60BR-20230524

Quant Time: May 30 00:04:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	257210	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	407219	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.410	117	377150	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	188781	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	172290	51.098	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 102.200%		
35) Dibromofluoromethane	5.282	113	137780	52.540	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 105.080%		
50) Toluene-d8	7.893	98	504516	50.579	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 101.160%		
62) 4-Bromofluorobenzene	10.629	95	174278	45.060	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 90.120%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.382	85	4975	1.733	ug/l	94
12) 1,1-Dichloroethene	2.575	96	868	0.314	ug/l	80
16) Acetone	2.626	43	5551	2.872	ug/l #	87
20) Methylene Chloride	3.044	84	809	0.232	ug/l #	85
21) trans-1,2-Dichloroethene	3.353	96	1955	0.660	ug/l #	81
27) cis-1,2-Dichloroethene	4.658	96	255257	77.467	ug/l	94
37) Ethyl Acetate	4.806	43	18465	3.678	ug/l #	82
44) Trichloroethene	6.536	130	925446	281.786	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052623\
Data File : VU054293.D
Acq On : 26 May 2023 18:12
Operator : JC/MD
Sample : 02936-15
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
60BR-20230524

Quant Time: May 30 00:04:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

