

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052722\
 Data File : VU048683.D
 Acq On : 27 May 2022 10:42
 Operator : SY/MD
 Sample : VU0527WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0527WBL01

Quant Time: May 28 00:15:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	216424	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	341869	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	352602	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	200763	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	166315	51.170	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.340%
35) Dibromofluoromethane	5.292	113	141997	58.220	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	116.440%
50) Toluene-d8	7.899	98	424892	48.599	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.200%
62) 4-Bromofluorobenzene	10.632	95	190209	57.526	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	115.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052722\
Data File : VU048683.D
Acq On : 27 May 2022 10:42
Operator : SY/MD
Sample : VU0527WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0527WBL01

Quant Time: May 28 00:15:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

