

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052722\
 Data File : VU048685.D
 Acq On : 27 May 2022 11:56
 Operator : SY/MD
 Sample : N3064-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TRIP-BLANK

Quant Time: May 28 00:16:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	197210	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	329140	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	341391	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	193663	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	148399	50.106	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.220%	
35) Dibromofluoromethane	5.292	113	126585	53.908	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.820%	
50) Toluene-d8	7.899	98	370681	44.038	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 88.080%#	
62) 4-Bromofluorobenzene	10.632	95	167589	52.645	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.300%	
Target Compounds						
16) Acetone	2.649	43	4820	3.193	ug/l	Qvalue # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052722\
Data File : VU048685.D
Acq On : 27 May 2022 11:56
Operator : SY/MD
Sample : N3064-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TRIP-BLANK

Quant Time: May 28 00:16:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

