

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052722\
 Data File : VU048687.D
 Acq On : 27 May 2022 12:43
 Operator : SY/MD
 Sample : N3064-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-1

Quant Time: May 28 00:17:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	194651	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	325906	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	339891	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.813	152	185687	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	168792	57.741	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	115.480%
35) Dibromofluoromethane	5.292	113	141076	60.676	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	121.360%
50) Toluene-d8	7.900	98	418373	50.197	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.400%
62) 4-Bromofluorobenzene	10.633	95	183377	58.177	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	116.360%
Target Compounds						
						Qvalue
16) Acetone	2.626	43	5429	3.644	ug/l	99
30) Chloroform	5.089	83	24561	5.522	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052722\
Data File : VU048687.D
Acq On : 27 May 2022 12:43
Operator : SY/MD
Sample : N3064-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-1

Quant Time: May 28 00:17:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

