

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052722\
 Data File : VU048688.D
 Acq On : 27 May 2022 13:07
 Operator : SY/MD
 Sample : N3064-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-2

Quant Time: May 28 00:17:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	193329	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	325490	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	338367	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	180648	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	160982	55.445	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.900%
35) Dibromofluoromethane	5.292	113	136471	58.770	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	117.540%
50) Toluene-d8	7.899	98	397529	47.757	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.520%
62) 4-Bromofluorobenzene	10.632	95	178524	56.709	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.420%
Target Compounds						
						Qvalue
16) Acetone	2.626	43	6379	4.311	ug/l	92
22) Diisopropyl ether	3.999	45	5014	0.713	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052722\
Data File : VU048688.D
Acq On : 27 May 2022 13:07
Operator : SY/MD
Sample : N3064-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-2

Quant Time: May 28 00:17:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

