

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052722\
 Data File : VU048689.D
 Acq On : 27 May 2022 13:30
 Operator : SY/MD
 Sample : N3064-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DUP-X

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/31/2022
 Supervised By : Mahesh Dadoda 05/31/2022

Quant Time: May 28 00:17:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	195824	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	320211	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	330956	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	182205	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	164793	56.035	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery	= 112.060%		
35) Dibromofluoromethane	5.292	113	140740	61.608	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 123.220%		
50) Toluene-d8	7.899	98	418957	51.161	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery	= 102.320%		
62) 4-Bromofluorobenzene	10.632	95	178546	57.651	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	= 115.300%		
Target Compounds						Qvalue
16) Acetone	2.629	43	7464	4.980	ug/l	98
22) Diisopropyl ether	4.002	45	4650m	0.653	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052722\
Data File : VU048689.D
Acq On : 27 May 2022 13:30
Operator : SY/MD
Sample : N3064-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP-X

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/31/2022
Supervised By :Mahesh Dadoda 05/31/2022

Quant Time: May 28 00:17:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

