

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052722\
 Data File : VU048700.D
 Acq On : 27 May 2022 17:49
 Operator : SY/MD
 Sample : N3064-06
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FIELD-BLANK

Quant Time: May 28 00:22:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 27 06:42:31 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	218740	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	355276	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	368928	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	207731	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	167705	51.051	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.100%
35) Dibromofluoromethane	5.292	113	143698	56.694	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	113.380%
50) Toluene-d8	7.899	98	429609	47.284	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.560%
62) 4-Bromofluorobenzene	10.632	95	193042	56.180	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.360%
Target Compounds						
					Qvalue	
16) Acetone	2.626	43	5597	3.343	ug/l	97
67) Ethyl Benzene	9.571	91	11871	0.902	ug/l	92
68) m/p-Xylenes	9.696	106	3624	0.704	ug/l	89
84) 1,2,4-Trimethylbenzene	11.468	105	13274	2.218	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052722\
Data File : VU048700.D
Acq On : 27 May 2022 17:49
Operator : SY/MD
Sample : N3064-06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FIELD-BLANK

Quant Time: May 28 00:22:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
Quant Title : SW846 8260
QLast Update : Fri May 27 06:42:31 2022
Response via : Initial Calibration

