


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052722\
 Data File : VU048708.D
 Acq On : 27 May 2022 21:20
 Operator : SY/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052622W.M
 Title : SW846 8260
 Last Update : Fri May 27 06:42:31 2022



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2876

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.0	13208	PASS
75	95	30	60	50.0	36739	PASS
95	95	100	100	100.0	73405	PASS
96	95	5	9	6.2	4534	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	79.4	58296	PASS
175	174	5	9	8.0	4667	PASS
176	174	95	101	98.2	57267	PASS
177	176	5	9	6.3	3613	PASS