

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052821\
 Data File : VU044004.D
 Acq On : 28 May 2021 13:19
 Operator : SY/MD
 Sample : M2250-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MDL-WATER-QT3-2021-06

Quant Time: May 29 02:05:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052621.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat May 29 02:01:06 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.226	114	14905	0.50	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	13692	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.759	152	6344	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.606	65	4199	0.46	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	92.00%
4) 1,2-Dichloroethane-d4	5.680	65	6173	0.45	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	90.00%
7) 1,2-Dichloropropane-d6	6.658	67	7077	0.54	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	108.00%
8) Toluene-d8	7.846	98	13675	0.50	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.00%
10) 1,1,2,2-Tetrachloroeth...	10.696	84	6915	0.55	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	110.00%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.606	62	448	0.03	ug/L #	35
6) Trichloroethene	6.508	95	390	0.03	ug/L	96
9) 1,2-Dibromoethane	8.866	107	272	0.03	ug/L #	90
12) 1,2,3-Trichloropropane	10.748	75	291	0.03	ug/L #	89
13) 1,2-Dibromo-3-chloropr...	12.924	75	65	0.04	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052821\
Data File : VU044004.D
Acq On : 28 May 2021 13:19
Operator : SY/MD
Sample : M2250-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-QT3-2021-06

Quant Time: May 29 02:05:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052621.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat May 29 02:01:06 2021
Response via : Initial Calibration

