

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052824\
 Data File : VU059023.D
 Acq On : 28 May 2024 12:24
 Operator : MD/SY
 Sample : P2570-04ME
 Misc : 4.97g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH6B2ME

Quant Time: May 29 02:41:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052324WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed May 29 02:38:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	80876	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	75365	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	27852	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	41792	48.858	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	97.720%	
7) Chloroethane-d5	1.875	69	39922	56.158	ug/L	-0.03
Spiked Amount	50.000	Range 70 - 130	Recovery	=	112.320%	
11) 1,1-Dichloroethene-d2	2.531	63	57633	44.101	ug/L	-0.03
Spiked Amount	50.000	Range 60 - 125	Recovery	=	88.200%	
21) 2-Butanone-d5	4.631	46	57925	98.586	ug/L	0.02
Spiked Amount	100.000	Range 40 - 130	Recovery	=	98.590%	
24) Chloroform-d	5.049	84	93071	63.435	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	126.860%#	
26) 1,2-Dichloroethane-d4	5.689	65	68094	75.888	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	151.780%#	
32) Benzene-d6	5.714	84	179061	69.691	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	139.380%#	
36) 1,2-Dichloropropane-d6	6.682	67	67057	79.016	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	158.040%#	
41) Toluene-d8	7.891	98	136757	59.381	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	118.760%	
43) trans-1,3-Dichloroprop...	8.174	79	21344	56.011	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	112.020%	
47) 2-Hexanone-d5	8.647	63	54654	155.715	ug/L	0.02
Spiked Amount	100.000	Range 45 - 130	Recovery	=	155.720%#	
56) 1,1,2,2-Tetrachloroeth...	10.753	84	98981	77.528	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	155.060%#	
66) 1,2-Dichlorobenzene-d4	12.190	152	41688	72.540	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	145.080%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052824\
Data File : VU059023.D
Acq On : 28 May 2024 12:24
Operator : MD/SY
Sample : P2570-04ME
Misc : 4.97g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH6B2ME

Quant Time: May 29 02:41:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052324WMA.M
Quant Title : VOC Analysis
QLast Update : Wed May 29 02:38:59 2024
Response via : Initial Calibration

