


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU052919\
 Data File : VU032345.D
 Acq On : 29 May 2019 08:30
 Operator : JC/SP
 Sample : BFB13
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB13

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR052219WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Fri May 24 02:55:50 2019



AutoFind: Scans 2865, 2866, 2867; Background Corrected with Scan 2852

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.9	12628	PASS
75	95	30	80	49.2	34792	PASS
95	95	100	100	100.0	70744	PASS
96	95	5	9	6.1	4309	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	93.5	66144	PASS
175	174	5	9	7.4	4864	PASS
176	174	95	101	96.8	64016	PASS
177	176	5	9	5.7	3679	PASS

Scan 2852 (10.260 min): VU032345.D
BFB13

m/z	Abundance
39.90	206.0
43.80	826.0
207.10	155.0

Instrument :
MSVOA_U
ClientSampleId :
BFB13

m/z	Abundance
36.00	516.0
37.10	2841.0
38.00	2573.0
38.90	1042.0
40.10	314.0
43.90	835.0
45.00	519.0
46.90	975.0
47.90	368.0
49.00	2161.0
50.00	12298.0
51.10	3768.0
51.90	168.0
54.90	293.0
56.10	804.0
57.00	2016.0
58.20	139.0
59.90	646.0
61.00	3340.0
62.10	2888.0
63.10	2507.0
63.90	570.0
65.10	967.0
68.00	7191.0
69.00	5906.0
70.00	581.0
72.00	624.0
73.10	2731.0
74.10	11443.0
75.10	35304.0
76.10	2874.0
77.00	508.0
78.00	342.0
78.90	2030.0
80.00	722.0
81.00	2181.0
81.70	450.0
86.00	204.0
87.00	2587.0
88.10	3092.0
90.90	319.0
92.00	1887.0
93.00	3307.0
94.10	8089.0
95.10	70800.0
96.00	4377.0
97.10	249.0
103.90	396.0
104.70	125.0
105.90	451.0
116.00	276.0
116.90	504.0
118.10	222.0
118.90	511.0
127.80	273.0
129.80	391.0
141.00	809.0
143.00	1395.0
143.80	127.0
145.10	596.0
145.90	130.0
147.90	109.0
149.90	101.0
151.90	117.0
157.10	162.0
171.20	205.0
172.00	925.0
174.00	63264.0
175.00	4718.0
176.00	60912.0
177.00	3682.0
178.10	190.0

Instrument :
MSVOA_U
ClientSampleId :
BFB13

m/z	Abundance
36.00	477.0
37.10	3185.0
38.00	2707.0
38.90	1149.0
39.90	281.0
43.90	895.0
45.00	484.0
47.00	933.0
48.00	372.0
49.00	2294.0
50.10	13088.0
51.10	3905.0
51.90	232.0
55.00	302.0
56.10	709.0
57.00	2273.0
58.20	212.0
59.90	663.0
61.00	3794.0
62.00	2640.0
63.10	2795.0
64.00	635.0
65.10	953.0
68.00	6807.0
69.10	6168.0
70.00	537.0
72.00	512.0
73.10	2763.0
74.10	11789.0
75.10	35872.0
76.10	2982.0
77.00	412.0
77.30	395.0
78.20	445.0
78.90	1990.0
79.90	714.0
81.00	2343.0
81.80	507.0
82.80	122.0
85.80	180.0
87.00	2839.0
88.00	2953.0
90.90	315.0
92.00	2046.0
93.10	3541.0
94.10	8660.0
95.10	73088.0
96.00	4392.0
97.20	220.0
103.90	374.0
104.70	101.0
105.70	426.0
115.90	228.0
117.00	442.0
118.10	245.0
118.90	600.0
127.90	362.0
129.80	408.0
134.00	115.0
136.60	108.0
141.00	783.0
143.00	1576.0
145.10	557.0
145.80	145.0
147.80	105.0
151.90	139.0
157.10	146.0
159.20	185.0
171.30	231.0
172.00	1044.0
174.00	68320.0
175.00	4992.0
176.00	66048.0
177.00	3798.0
177.80	159.0
178.10	167.0

Instrument :
MSVOA_U
ClientSampleId :
BFB13

m/z	Abundance
36.00	404.0
37.00	3198.0
37.90	2533.0
39.00	1184.0
39.90	343.0
43.90	761.0
44.90	417.0
47.00	712.0
48.00	394.0
49.00	2177.0
50.00	12500.0
51.10	3788.0
51.90	242.0
55.00	293.0
56.10	576.0
57.00	1990.0
58.20	235.0
59.90	653.0
61.10	3652.0
62.00	2557.0
63.10	2919.0
64.00	557.0
65.10	922.0
68.00	6243.0
69.10	6014.0
70.10	470.0
71.90	341.0
73.10	2792.0
74.00	11304.0
75.10	33200.0
76.10	2604.0
77.30	489.0
78.10	507.0
79.00	1869.0
79.90	656.0
81.00	2174.0
81.90	577.0
82.80	116.0
85.80	115.0
87.00	2736.0
87.90	2749.0
90.90	242.0
92.00	2016.0
93.10	3414.0
94.10	8370.0
95.10	68344.0
96.00	4158.0
97.20	180.0
103.80	245.0
105.70	336.0
115.90	165.0
117.00	405.0
118.20	195.0
119.00	623.0
127.90	406.0
129.80	363.0
135.00	112.0
141.00	743.0
143.00	1349.0
145.00	518.0
145.80	122.0
147.80	125.0
151.90	120.0
156.80	143.0
159.20	272.0
171.20	263.0
172.00	1155.0
174.00	66848.0
175.00	4882.0
176.00	65088.0
177.00	3557.0

Instrument :
MSVOA_U
ClientSampleId :
BFB13