


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052924\
Data File : VU059044.D
Acq On : 29 May 2024 10:28
Operator : MD/SY
Sample : BFB055
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB055

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Title : VOC Analysis
Last Update : Sat Jun 01 01:17:21 2024



AutoFind: Scans 2904, 2905, 2906; Background Corrected with Scan 2893

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.4	4421	PASS
75	95	30	80	53.5	11620	PASS
95	95	100	100	100.0	21717	PASS
96	95	5	9	7.2	1567	PASS
173	174	0.00	2	1.6	270	PASS
174	95	50	120	77.8	16906	PASS
175	174	5	9	6.9	1166	PASS
176	174	95	101	95.1	16078	PASS
177	176	5	9	6.7	1075	PASS

m/z	Abundance
39.80	152.0
43.80	265.0

Instrument :
MSVOA_U
ClientSampleId :
BFB055

m/z	Abundance
35.90	159.0
36.90	1045.0
37.90	1119.0
39.00	480.0
39.80	169.0
43.90	419.0
44.80	222.0
46.80	254.0
47.90	137.0
48.90	1043.0
49.90	4441.0
50.90	1412.0
55.90	483.0
56.90	747.0
59.90	163.0
60.90	1226.0
61.90	1317.0
62.90	974.0
63.90	179.0
64.90	257.0
67.90	2472.0
68.90	2553.0
69.90	212.0
71.70	123.0
72.90	941.0
73.90	4102.0
74.90	11322.0
75.90	994.0
76.90	220.0
78.80	815.0
79.90	271.0
80.80	995.0
81.80	396.0
86.90	979.0
87.80	1027.0
91.90	585.0
93.00	1093.0
93.90	2290.0
94.90	20896.0
95.90	1579.0
105.90	130.0
116.80	175.0
118.90	182.0
140.70	186.0
142.80	307.0
171.80	238.0
172.90	249.0
173.90	15853.0
174.90	1095.0
175.90	15322.0
176.90	1048.0

Instrument :
MSVOA_U
ClientSampleId :
BFB055

m/z	Abundance
35.90	203.0
36.90	1139.0
37.90	1087.0
38.90	534.0
39.80	224.0
43.90	399.0
44.90	228.0
46.80	251.0
48.00	141.0
48.90	1049.0
49.90	4552.0
50.90	1435.0
55.90	486.0
56.90	719.0
59.80	209.0
60.90	1364.0
61.90	1250.0
62.90	1069.0
63.90	232.0
64.80	306.0
67.90	2454.0
68.90	2657.0
69.90	184.0
71.60	136.0
73.00	1021.0
73.90	4354.0
74.90	11983.0
75.90	901.0
76.90	261.0
77.90	143.0
78.80	956.0
79.90	273.0
80.80	1031.0
81.80	363.0
86.90	1085.0
87.80	1188.0
91.90	526.0
93.00	1129.0
93.90	2472.0
95.00	22248.0
96.00	1610.0
105.90	134.0
116.80	162.0
118.90	171.0
140.80	185.0
142.80	353.0
171.90	303.0
172.90	277.0
173.90	17392.0
174.90	1197.0
175.90	16704.0
176.90	1104.0

Instrument :
MSVOA_U
ClientSampleId :
BFB055

m/z	Abundance
35.90	235.0
36.90	1150.0
37.90	1002.0
38.90	511.0
39.80	278.0
43.90	384.0
44.90	227.0
46.90	254.0
48.00	120.0
48.90	978.0
49.90	4271.0
50.90	1353.0
55.90	398.0
56.90	580.0
59.80	231.0
61.00	1287.0
61.90	1087.0
62.90	1078.0
64.00	252.0
64.80	302.0
67.90	2340.0
69.00	2530.0
69.90	214.0
72.00	126.0
73.00	1108.0
73.90	4081.0
74.90	11555.0
75.90	855.0
76.90	263.0
77.90	137.0
78.90	1038.0
79.90	233.0
80.90	941.0
81.80	308.0
86.90	1006.0
87.90	1122.0
91.90	548.0
93.00	1178.0
93.90	2582.0
95.00	22008.0
96.00	1512.0
105.80	136.0
116.80	139.0
118.80	142.0
140.80	193.0
142.80	349.0
171.90	329.0
172.80	284.0
173.90	17472.0
175.00	1205.0
175.90	16207.0
176.90	1074.0

Instrument :
MSVOA_U
ClientSampleId :
BFB055