

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU053019\
 Data File : VU032372.D
 Acq On : 30 May 2019 11:31
 Operator : JC/SP
 Sample : VU0530MBL01
 Misc : 5.00g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK44

Quant Time: May 31 04:47:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM052819WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 31 04:34:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	773478	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	768124	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	325229	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	247658	43.44	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.88%
7) Chloroethane-d5	1.68	69	208086	45.20	ug/L	0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.40%
11) 1,1-Dichloroethene-d2	2.27	63	346680	34.23	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.46%
21) 2-Butanone-d5	4.17	46	330676	88.13	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.13%
24) Chloroform-d	4.64	84	447508	44.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.70%
26) 1,2-Dichloroethane-d4	5.31	65	283137	47.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.90%
32) Benzene-d6	5.34	84	953669	47.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.68%
36) 1,2-Dichloropropane-d6	6.32	67	294800	47.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.28%
41) Toluene-d8	7.56	98	854930	45.48	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.96%
43) trans-1,3-Dichloropropene-	7.85	79	127627	44.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.14%
47) 2-Hexanone-d5	8.31	63	229810	87.59	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.59%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	425950	44.58	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.16%
64) 1,2-Dichlorobenzene-d4	11.86	152	334318	50.41	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.82%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU053019\
Data File : VU032372.D
Acq On : 30 May 2019 11:31
Operator : JC/SP
Sample : VU0530MBL01
Misc : 5.00g/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK44

Quant Time: May 31 04:47:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM052819WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 31 04:34:14 2019
Response via : Initial Calibration

