

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU053020\
 Data File : VU038400.D
 Acq On : 30 May 2020 00:38
 Operator : JC/MD
 Sample : L2766-05DL 200X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BE4Y8DL

Quant Time: May 30 03:16:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR050720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat May 30 00:03:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	302649	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	294781	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	144109	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	90259	3.26	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	65.20%
7) Chloroethane-d5	1.93	69	84507	3.53	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	70.60%
11) 1,1-Dichloroethene-d2	2.59	63	138954	3.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.00%
20) 2-Butanone-d5	4.67	46	352708	51.64	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.28%
24) Chloroform-d	5.10	84	197207	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.80%
26) 1,2-Dichloroethane-d4	5.74	65	120874	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.40%
32) Benzene-d6	5.76	84	410738	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.00%
36) 1,2-Dichloropropane-d6	6.72	67	129217	4.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.60%
41) Toluene-d8	7.92	98	348403	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.80%
43) trans-1,3-Dichloropropene-	8.20	79	49977	4.38	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.60%
46) 2-Hexanone-d5	8.65	63	299368	55.51	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.02%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	109141	4.85	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.00%
64) 1,2-Dichlorobenzene-d4	12.20	152	141421	5.04	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.80%

Target Compounds

					Qvalue
8) Chloroethane	1.95	64	4825	0.24 ug/L	# 75
33) Benzene	5.81	78	542727	5.75 ug/L	100
42) Toluene	7.99	91	893137	8.96 ug/L	98
51) Chlorobenzene	9.46	112	35603	0.58 ug/L	93
52) Ethylbenzene	9.59	91	116717	1.08 ug/L	99
53) m,p-Xylene	9.71	106	104770	2.53 ug/L	96
54) o-Xylene	10.11	106	35549	0.89 ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU053020\
Data File : VU038400.D
Acq On : 30 May 2020 00:38
Operator : JC/MD
Sample : L2766-05DL 200X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BE4Y8DL

Quant Time: May 30 03:16:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR050720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat May 30 00:03:10 2020
Response via : Initial Calibration

