

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053023\
 Data File : VU054303.D
 Acq On : 30 May 2023 10:38
 Operator : JC/MD
 Sample : VU0530WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0530WBL01

Quant Time: May 31 06:22:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.366	168	250385	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.244	114	415891	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.411	117	393028	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	192425	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	171376	52.212	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.420%
35) Dibromofluoromethane	5.282	113	139836	52.212	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.420%
50) Toluene-d8	7.893	98	513655	50.422	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.840%
62) 4-Bromofluorobenzene	10.626	95	179591	45.466	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053023\
Data File : VU054303.D
Acq On : 30 May 2023 10:38
Operator : JC/MD
Sample : VU0530WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0530WBL01

Quant Time: May 31 06:22:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

