

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053023\
 Data File : VU054306.D
 Acq On : 30 May 2023 12:12
 Operator : JC/MD
 Sample : 02912-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GM38-RW3-MW1-20230523

Quant Time: May 31 06:24:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.366	168	255495	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	423236	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.410	117	409650	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	206447	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	173880	51.916	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.840%	
35) Dibromofluoromethane	5.282	113	143895	52.795	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.580%	
50) Toluene-d8	7.893	98	529473	51.072	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.140%	
62) 4-Bromofluorobenzene	10.626	95	193051	48.025	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.060%	
Target Compounds						Qvalue
44) Trichloroethene	6.536	130	43816	12.837	ug/l	90
64) Tetrachloroethene	8.549	164	3448	1.118	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053023\
Data File : VU054306.D
Acq On : 30 May 2023 12:12
Operator : JC/MD
Sample : 02912-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GM38-RW3-MW1-20230523

Quant Time: May 31 06:24:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

