

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053023\
 Data File : VU054307.D
 Acq On : 30 May 2023 12:36
 Operator : JC/MD
 Sample : 02912-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BPOW1-7-20230523

Quant Time: May 31 06:24:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	254991	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	423148	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.414	117	397735	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	203037	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.697	65	172737	51.676	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.360%
35) Dibromofluoromethane	5.282	113	138544	50.842	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.680%
50) Toluene-d8	7.893	98	520213	50.190	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.380%
62) 4-Bromofluorobenzene	10.629	95	185801	46.231	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.460%
Target Compounds						Qvalue
16) Acetone	2.642	43	2044	1.067	ug/l	98
20) Methylene Chloride	3.047	84	1112	0.322	ug/l #	90
44) Trichloroethene	6.546	130	892	0.261	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053023\
Data File : VU054307.D
Acq On : 30 May 2023 12:36
Operator : JC/MD
Sample : 02912-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BPOW1-7-20230523

Quant Time: May 31 06:24:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

