

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053023\
 Data File : VU054309.D
 Acq On : 30 May 2023 13:23
 Operator : JC/MD
 Sample : 02912-05DL 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE103D1-20230523DL

Quant Time: May 31 06:25:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	259100	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	423210	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.414	117	403634	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	202587	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.697	65	175063	51.542	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.080%
35) Dibromofluoromethane	5.282	113	142612	52.327	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.660%
50) Toluene-d8	7.893	98	525924	50.733	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.460%
62) 4-Bromofluorobenzene	10.629	95	188320	46.851	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.700%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.578	101	1507	0.560	ug/l #	91
20) Methylene Chloride	3.041	84	6755	1.923	ug/l	92
44) Trichloroethene	6.536	130	153834	45.071	ug/l	89
64) Tetrachloroethene	8.549	164	1114	0.367	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053023\
Data File : VU054309.D
Acq On : 30 May 2023 13:23
Operator : JC/MD
Sample : 02912-05DL 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE103D1-20230523DL

Quant Time: May 31 06:25:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

