

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053023\
 Data File : VU054310.D
 Acq On : 30 May 2023 13:47
 Operator : JC/MD
 Sample : 02912-06
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BPOW1-8-20230523

Quant Time: May 31 06:25:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	257565	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.244	114	417065	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.414	117	400493	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	200437	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	175703	52.038	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.080%	
35) Dibromofluoromethane	5.282	113	139750	52.033	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.060%	
50) Toluene-d8	7.893	98	517939	50.699	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.400%	
62) 4-Bromofluorobenzene	10.626	95	185938	46.940	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.880%	
Target Compounds						
					Qvalue	
16) Acetone	2.639	43	1804	0.932	ug/l	93
20) Methylene Chloride	3.041	84	1206	0.345	ug/l #	85
44) Trichloroethene	6.539	130	6481	1.927	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053023\
Data File : VU054310.D
Acq On : 30 May 2023 13:47
Operator : JC/MD
Sample : 02912-06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BPOW1-8-20230523

Quant Time: May 31 06:25:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

