

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053023\
 Data File : VU054315.D
 Acq On : 30 May 2023 15:45
 Operator : JC/MD
 Sample : 02912-09
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE103D3-20230523

Quant Time: May 31 06:27:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.366	168	260121	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	426180	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.410	117	396117	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	189531	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	172198	50.499	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.000%
35) Dibromofluoromethane	5.282	113	142149	51.794	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.580%
50) Toluene-d8	7.893	98	524947	50.286	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.580%
62) 4-Bromofluorobenzene	10.629	95	180895	44.690	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.380%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.571	101	2291	0.848	ug/l	87
16) Acetone	2.633	43	1480	0.757	ug/l #	85
20) Methylene Chloride	3.038	84	973	0.276	ug/l #	91
44) Trichloroethene	6.536	130	329428	95.844	ug/l	92
64) Tetrachloroethene	8.546	164	773	0.259	ug/l #	88
96) 1,2,3-Trichlorobenzene	14.336	180	200	1.300	ug/l #	50

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053023\
Data File : VU054315.D
Acq On : 30 May 2023 15:45
Operator : JC/MD
Sample : 02912-09
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE103D3-20230523

Quant Time: May 31 06:27:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

