

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053023\
 Data File : VU054317.D
 Acq On : 30 May 2023 16:33
 Operator : JC/MD
 Sample : 02912-10
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EB02-20230523

Quant Time: May 31 06:28:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	251833	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	417088	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.411	117	394771	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.806	152	185008	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	170062	51.514	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 103.020%		
35) Dibromofluoromethane	5.282	113	138350	51.509	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.020%		
50) Toluene-d8	7.893	98	519591	50.858	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 101.720%		
62) 4-Bromofluorobenzene	10.629	95	178702	45.111	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 90.220%		
Target Compounds						
					Qvalue	
16) Acetone	2.633	43	3538	1.870	ug/l	95
20) Methylene Chloride	3.041	84	1046	0.306	ug/l #	81
30) Chloroform	5.089	83	5450	1.006	ug/l	98
44) Trichloroethene	6.539	130	4267	1.269	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053023\
Data File : VU054317.D
Acq On : 30 May 2023 16:33
Operator : JC/MD
Sample : 02912-10
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EB02-20230523

Quant Time: May 31 06:28:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

