

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053023\
 Data File : VU054325.D
 Acq On : 30 May 2023 19:43
 Operator : JC/MD
 Sample : 02912-13DL 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE103D2-20230523DL

Quant Time: May 31 06:31:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 26 06:40:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	246703	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.243	114	411189	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.410	117	388953	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.809	152	186275	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.694	65	171508	53.032	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.060%	
35) Dibromofluoromethane	5.282	113	140156	52.930	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.860%	
50) Toluene-d8	7.893	98	510677	50.703	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.400%	
62) 4-Bromofluorobenzene	10.629	95	175701	44.990	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.980%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.571	101	1331	0.519	ug/l	# 85
20) Methylene Chloride	3.044	84	6741	2.016	ug/l	96
44) Trichloroethene	6.536	130	103216	31.124	ug/l	91
64) Tetrachloroethene	8.549	164	633	0.216	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053023\
Data File : VU054325.D
Acq On : 30 May 2023 19:43
Operator : JC/MD
Sample : 02912-13DL 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE103D2-20230523DL

Quant Time: May 31 06:31:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U052523W.M
Quant Title : SW846 8260
QLast Update : Fri May 26 06:40:51 2023
Response via : Initial Calibration

