

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053024\
 Data File : VU059091.D
 Acq On : 30 May 2024 14:03
 Operator : MD/SY
 Sample : P2589-19ME
 Misc : 5.45g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH6D3ME

Quant Time: May 31 03:22:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 31 03:18:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.239	114	173729	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	177664	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	81873	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	70512	67.788	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	135.580%#
7) Chloroethane-d5	1.875	69	62354	69.979	ug/L	-0.03
Spiked Amount	50.000	Range	70 - 130	Recovery	=	139.960%#
11) 1,1-Dichloroethene-d2	2.544	63	94452	46.002	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.000%
21) 2-Butanone-d5	4.615	46	117970	120.323	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	120.320%
24) Chloroform-d	5.049	84	133324	64.385	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	128.760%#
26) 1,2-Dichloroethane-d4	5.689	65	91087	71.769	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	143.540%#
32) Benzene-d6	5.715	84	273082	66.970	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	133.940%#
36) 1,2-Dichloropropane-d6	6.682	67	92353	65.686	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	131.380%#
41) Toluene-d8	7.891	98	236047	68.191	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	136.380%#
43) trans-1,3-Dichloroprop...	8.174	79	39894	70.197	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	140.400%#
47) 2-Hexanone-d5	8.637	63	73136	112.489	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	112.490%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	142548	59.907	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	119.820%
66) 1,2-Dichlorobenzene-d4	12.190	152	87719	71.794	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	143.580%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053024\
Data File : VU059091.D
Acq On : 30 May 2024 14:03
Operator : MD/SY
Sample : P2589-19ME
Misc : 5.45g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH6D3ME

Quant Time: May 31 03:22:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 31 03:18:55 2024
Response via : Initial Calibration

