

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053024\
 Data File : VU059092.D
 Acq On : 30 May 2024 14:27
 Operator : MD/SY
 Sample : P2623-01ME
 Misc : 5.12g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHBQ8ME

Quant Time: May 31 03:22:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 31 03:18:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	204921	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	213659	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	97499	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	62045	50.568	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.140%
7) Chloroethane-d5	1.875	69	55391	52.702	ug/L	-0.03
Spiked Amount	50.000	Range	70 - 130	Recovery	=	105.400%
11) 1,1-Dichloroethene-d2	2.544	63	83999	34.684	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.360%
21) 2-Butanone-d5	4.618	46	107127	92.632	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.630%
24) Chloroform-d	5.049	84	122950	50.337	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.680%
26) 1,2-Dichloroethane-d4	5.689	65	83147	55.541	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	111.080%
32) Benzene-d6	5.714	84	252816	51.555	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.120%
36) 1,2-Dichloropropane-d6	6.682	67	84486	49.968	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.940%
41) Toluene-d8	7.891	98	213086	51.187	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.380%
43) trans-1,3-Dichloroprop...	8.174	79	36554	53.484	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	106.960%
47) 2-Hexanone-d5	8.637	63	63912	81.741	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.740%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	129590	45.286	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.580%
66) 1,2-Dichlorobenzene-d4	12.190	152	75938	52.191	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.380%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU053024\
Data File : VU059092.D
Acq On : 30 May 2024 14:27
Operator : MD/SY
Sample : P2623-01ME
Misc : 5.12g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHBQ8ME

Quant Time: May 31 03:22:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM052924WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 31 03:18:55 2024
Response via : Initial Calibration

